



Medicines & Healthcare products
Regulatory Agency

Traceability – regulatory aspects

GS1 Healthcare Conference – 10 April 2019

*Andy Crosbie – Manager – Post-market Surveillance Strategy
Medicines and Healthcare Products Regulatory Agency*




Key topics

- Overview of new Medical Device Regulations
- Traceability requirements
- Supply chain – recalls – implant cards

New medical device regulations

Two EU Regulations



**MDR - Regulation (EU) 2017/745 –
published 5 May 2017**

**Medical Devices
Regulations**

**in-Vitro Diagnostic
Devices Regulations**

**IVDR - Regulation (EU) 2017/746
– published 5 May 2017**

**UDI assignment and submission of UDI core data elements to the European database by 26
May 2020 (medical devices) or 26 May 2022 (IVDs)**

New European Regulations: major upgrade



UDI is one small(!) part of a complete revamp of the European Medical Device (MDR) and IVD (IVDR) regulations

Council Directive
93/42/EEC - 14
June 1993

60 pages

23 Articles

12 Annexes

Regulation (EU)
2017/745 – 5 May
2017

175 pages

123 Articles

17 Annexes

UDIs enhance traceability and improve post-market safety

Using a Unique Device Identification (UDI) system **based on international guidance** should significantly enhance the post-market safety of medical devices by:

improving incident reporting

better targeting of recalls

better monitoring by competent authorities

reducing medical errors

fighting against counterfeit devices

improving purchase-policy and stock-management by hospitals

Regulation (EU) 2017/745: Recital 41



UDI with a GTIN, Expiry and Serial Number

GS1 = UDI

(though other UDI systems may also be acceptable!)

Translation UDI to GS1

UDI Unique Device Identification	GS1 Standards Product Identification
UDID Data Elements linked to the Device Identifier	GDSN Attributes mapped to each UDID data element
DI = Device Identifier (DI)	GTIN Global Trade Item Number
<i>Production data is not stored in UDI or GDSN databases</i>	
PI = Production Identifier (PI) (if applicable) Production Identifier data will vary by medical device type and manufacturer current practice.	AI Application Identifiers (AI) <ul style="list-style-type: none">• Expiration Date AI(17) e.g. 141120• Lot/Batch AI(10) e.g. 1234AB• Serial Number AI(21) e.g. 12345XYZ
DI + PI = UDI	GTIN -or- GTIN + AI(s) = UDI

Requirements of the European UDI system

UDI system should allow the **identification and traceability** of medical devices (excludes custom-made and clinical trial devices) and IVDs

UDIs should be in **two parts**:

- **device identifier** (UDI-DI) – *static*
- **production identifier** (UDI-PI) – *dynamic*

UDIs should be **placed on device labels** (machine readable + plain text)

UDIs should be **stored by manufacturers, importers, distributors and hospitals** (using electronic means)

Centralised **European UDI database** (part of Eudamed)

How UDIs should be used

UDI should be placed on the labels of devices and must:


be used for reporting serious incidents and field safety corrective actions (recalls)

be referenced in information provided to patients who have received implants (implant cards or electronic)

be included in technical / regulatory documentation

Traceability throughout the supply chain

Manufacturer → importer/distributor → hospital → patient




Recording of UDIs

Manufacturers, importers, distributors and hospitals must store and keep UDIs (both device identifier and production identifiers) by electronic means



SCAN  SAFETY
Patient, Product, Place, Process.

Using UDI / GS1 standards improves patient safety – implant tracking and surveillance



MHRA Medical Device Alerts



Medical Device Alert

MDA/2015/024 Issued: 25 June 2015 at 14:00

Metal-on-metal (MoM) hip replacements:
Birmingham Hip™ Resurfacing (BHR) system (Smith & Nephew Orthopaedics)

Summary

Higher than expected revision rate for certain patient groups implanted with the Birmingham Hip Resurfacing (BHR) system. Guidance provided on implantation and patient management.

Action

- Do not implant BHR devices in:
 - > female patients
 - > patients requiring femoral heads sized 46mm or smaller.
- Only use 48mm BHR heads in the specific circumstance of intra-operative downsizing from a pre-operatively templated 50mm to a measured 48mm at the time of surgery.
- Return all unused BHR femoral heads sized 46mm and smaller and their corresponding acetabular and dysplasia cups to the manufacturer.
- Follow up patients implanted with BHR hips that fall within the scope of this Medical Device Alert i.e.:
 - > all symptomatic patients
 - > all female patients
 - > all patients implanted with head sizes 46mm or smaller

in line with recommendations in the table below (based on the advice given in MDA/2012/036)

	Management recommendations for patients implanted with BHR hips that fall within the scope of this Medical Device Alert
Patient follow-up	Annually for the life of the implant
Imaging: MARS MRI or ultrasound	Recommended in all cases
1 st blood metal ion level test	Yes
Results of 1 st blood metal ion level test	Blood metal ion level >7ppb indicates potential for soft tissue reaction
2 nd blood metal ion level test	Yes - 3 months after 1 st blood test if result was >7ppb
Results of 2 nd blood metal ion level test	Blood metal ion level >7ppb indicates potential for soft tissue reaction especially if greater than previously
Consider need for revision	If imaging is abnormal and/or blood metal ion levels rising

Table footnotes:

- Blood metal ion testing to be in whole blood
- 7 parts per billion (ppb) equals 119 nmol/L cobalt or 134.5 nmol/L chromium

Device details

Catalogue number	Unique Device Identifier (GS1)
BHR head	
74121138	03596010502766
74123140	03596010552402
74121142	03596010502773
74123144	03596010552419
74121146	03596010502780
BHR acetabular cup	
74120144	03596010502537
74120146	03596010502544
74122146	03596010565792
74122148	03596010552266
74120148	03596010502551
74120150	03596010502568
74122150	03596010552273
74122152	03596010552280
74120152	03596010502575
74120154	03596010502582
BHR dysplasia cup	
74120246	03596010502650
74122248	03596010552358
74120250	03596010502667
74122252	03596010552365
74120254	03596010502674



Field Safety Corrective Actions and Recalls
















Manufacturers should:


- Reference UDIs in their Field Safety Notices (FSNs)
- Send customers and distributors electronic lists of all products affected by a FSCA in an Excel spreadsheet

“ These recommendations give healthcare providers everything they need to implement more automated and secure GS1-based FSCA processes, in line with the Department of Health eProcurement strategy.”

- Jackie Pomroy, Head of Supply Chain at NHS South of England Procurement Services

European implant cards

Implant Card	Life-saver 5000 Cardiac Pacemaker	Life-electrode 123 Cardiac Pacemaker electrode
 <u>John</u>	 <u>Lifesaver 5000</u> <u>Cardiac Pacemaker</u>	 <u>Life-electrode 123</u> <u>Cardiac Pacemaker electrode</u>
<u>Doe</u>	SN <u>VDD 0123456789</u>	SN <u>E123 9876543210</u>
 <u>XYZ Healthcare Center</u>	LOT	LOT
<u>Streetname 123, City</u>	UDI  <u>(01) 2 081019001 002 4</u>  <u>(17)080100(10)1Q34</u>	UDI  <u>(01) 2 081019001 002 4</u>  <u>(17)080100(10)1Q34</u>
 <u>01.01.2021</u>	 <u>Lifesaver Co. & KG</u> <u>Street 123, 12345 City</u>	 <u>Lifesaver Co. & KG</u> <u>Street 123, 12345 City</u>
	 <u>www.lifesaver.eu/manual</u>	 <u>www.lifesaver.eu/manual</u>



(01) 0 0643169 00176 3 (17) 160712 (21) 21A11F4855

UDI with a GTIN, Expiry and Serial Number

Thank you

andy.crosbie@mhra.gov.uk